
TWL Calibration Data Summary

Serial Number: 0001733

Full-Scale Flow Rate: 80
(yields 20mA output)

Zero Flow Rate: (yields 4mA output)

Bridge Resistance: 350.7 ohms

Full-Scale Bridge Output: 2.0095 mV/V

Calibration Resistor (R7): 47332 ohms

Calibration Voltage: 0.9184 volts

Calibration Flow Rate: 76.66

Calibration Current: 19.33 mA

Zero Switch Setting: 0

TB3 Jumper Position: + (POSITIVE)

Span Switch Setting - Undamped: 0

Span Switch Setting - Damped: 8

Additional Information:
